



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017  
& ANSI/NCSL Z540-1-1994

THERMOTRON INDUSTRIES,  
A VENTUREDYNE, LTD., COMPANY  
291 Kollen Park Drive  
Holland, MI 49423  
Gray H. Warner Phone: 616 392 6550

CALIBRATION

Valid To: September 30, 2026

Certificate Number: 1917.01

In recognition of the successful completion of the A2LA evaluation process (including an assessment of the organization's compliance with A2LA's Calibration Program Requirements), accreditation is granted to this laboratory to perform the following calibrations<sup>1, 6</sup>:

I. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC <sup>2, 5</sup> (±)	Comments
DC Voltage – Generate <sup>3</sup>	(0 to 200) mV 200 mV to 2 V (2 to 20) V	0.015 % + 0.58 mV 0.015 % + 0.63 mV 0.015 % + 3.8 mV	Transcat 23894E
DC Voltage – Measure <sup>3</sup>	(0 to 10) V	35 µV/V + 3.5 mV	Agilent 34401A
DC Current – Generate <sup>3</sup>	(0 to 24) mA	0.015 % + 9.7 µA	Transcat 23894E
DC Current – Measure <sup>3</sup>	(0 to 100) mA 100 mA to 1 A	0.1 % + 0.58 mA 0.1 % + 0.63 mA	Agilent 34401A

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Electrical Simulation of Thermocouple Indicating Systems <sup>3</sup> –			
Type J	(-210 to -47) °C (-47 to 760) °C	0.68 °C 0.4 °C	Omega CL-27 or Transcat 23700T
Type K	(-200 to -46) °C (-46 to 1372) °C	0.68 °C 0.4 °C	
Type T	(-200 to -47) °C (-47 to 399) °C	0.68 °C 0.4 °C	
Electrical Simulation of RTD Indicating Systems <sup>3</sup> –			
100 Ω	(-200 to 800) °C	0.4 °C	Fluke 712 or Omega CL-27
1000 Ω	(-200 to 630) °C	0.4 °C	

Parameter/Range	Frequency	CMC <sup>2,5</sup> (±)	Comments
Thermotron PTS Only AC Voltage – Measure <sup>3</sup>			
(0 to 1) V	10 Hz to 20 kHz	0.06 % + 4.9 mV	Agilent 34401A
(1 to 10) V	10 Hz to 20 kHz	0.06 % + 3.5 mV	
AC Voltage – Measure <sup>3</sup>			
(0 to 3) V	(20 to 50) Hz 50 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz	1 % + 7.6 mV 0.2 % + 1.9 mV 0.5 % + 4.0 mV 2 % + 18 mV	Fluke 45
(3 to 30) V	(20 to 50) Hz 50 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz	1 % + 76 mV 0.2 % + 19 mV 0.5 % + 40 mV 2.4 % + 160 mV	

## II. Thermodynamics

Parameter/Equipment	Range	CMC <sup>2,7</sup> (±)	Comments
Temperature – Measure <sup>3</sup>	(-80 to -46) °C (-46 to 260) °C	0.75 °C 0.52 °C	Omega CL-27 or Transcat 23700T with type T thermocouple
Relative Humidity – Measuring Equipment <sup>3</sup>	11.3 % RH 33.1 % RH 75.5 % RH 97.6 % RH	1.4 % RH 1.3 % RH 1.6 % RH 2.1 % RH	Vaisala HMK-15 with:  LiCl MgCl <sub>2</sub> NaCl K <sub>2</sub> SO <sub>4</sub>
Relative Humidity – Measure <sup>3</sup>	(10 to 90) % RH (90 to 95) % RH	1.8 % RH 2.9 % RH	Vaisala HM -70 with HMP-77 humidity probe

## III. Time & Frequency

Parameter/Equipment	Frequency	CMC <sup>2,4,7</sup> (±)	Comments
Frequency – Measure <sup>3</sup>	(40 to 100) Hz 100 Hz to 100 kHz	0.01 % + 0.066 Hz 0.01 % + 35 Hz	Agilent 34401A

<sup>1</sup> This laboratory offers commercial calibration service and field calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k=2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

<sup>4</sup> In the statement of CMC, a percent represents a % of reading unless otherwise noted.

<sup>5</sup> The measurands stated are generated using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure the measurand in the ranges indicated. CMCs are expressed as either a specific value that covers the full range or as a fraction of the reading plus a fixed floor specification.

<sup>6</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.

<sup>7</sup> The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.



# Accredited Laboratory

A2LA has accredited

## **THERMOTRON INDUSTRIES, A VENTUREDYNE, LTD., COMPANY**

Holland, MI

for technical competence in the field of

### Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and the requirements of ANSI/NCSL Z540.3-2006 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 26<sup>th</sup> day of August 2024.

A blue ink signature of Trace McInturff.

Mr. Trace McInturff Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 1917.01  
Valid to September 30, 2026

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.